


INOGEN ONE[®] G5[™] System



Inogen One
AC Power Supply*
Item #BA-501



Inogen One G5
DC Power Cable*
Item #BA-306



Inogen One G5
Lithium Ion Battery*
Up to 6.5 hours run time
Item #BA-500



Inogen One G5
Carry Bag*
Item #CA-500



Inogen One[®]G5
Extended Life Lithium Ion Battery**
Up to 13 hours run time
Item #BA-516

The Inogen One G5 System (IS-500):






- Compact and Lightweight
- Flow Settings 1-6
- Intelligent Pulse Oxygen Delivery
- Designed for 24 Hour Use
- Easy-to-read LCD display
- Inogen Connect App

*Included with system

**Accessory sold separately



PRODUCT COMPARISON

	Inogen G5	SimplyGo Mini	Zen-O Lite	iGO2	Precision Live Active 5
					
Settings	1-6 Pulse	1 - 5 Pulse	1 - 5 Pulse	1 - 5 Pulse	1 - 5 Pulse
Weight	2.13 Kgs	2.26 Kgs	2.4. Kgs	2.2 Kgs	2.2 Kgs
Sound	38 dbA @ setting 2	42 dBA @ setting 2	37 dbA @ setting 2	37 dbA @ setting 2	43 dbA @ setting 2
Maximum Oxygen Output	1,250 mL/min	1,000 mL/min	1,050 mL/min	1,014 mL/min	1,000 mL/min
Battery life at setting 2	Up to 12 hours	Up to 4.5 hours Extended: Up to 9 hours	Up to 4 hours	Up to 3.5 hours	Up to 4.3 hours
Removable Battery Access	Two battery sizes, slides onto bottom of device	Two battery sizes, slides longways onto bottom of device	Compartment at bottom of device	One rechargeable battery	One Battery at the side compartment
Patient Replaceable Sieve Beds	Yes	No, must be sent in for service	Yes	No, must be sent in for service	Yes
Maximum Cannula Length	25 ft.	10 ft.	25 ft.	25 ft.	4 ft.
No Breath Detect (NBD) Response	Audible and visual alarm if NBD for 60 sec. pulses at 17 BPM until another breath is detected	Audible and visual alarm if NBD for 2 minutes, pulses at 12 BPM until another breath is detected. Unit shuts down after 15 min. of autopulsing. Upon start-up can take up to 12 min. to activate.	Audible and visual alarm after 15 sec. NBD and until another breath is detected, unit does not autopulse, must switch to another source of oxygen	Audible and visual alarm if NBD after 60 seconds. Unit will go into AutoBreath mode until breathing is detected again.	Visual Alarm after 60 sec.NBD and unit will auto pulse at 20 pulses per minute until breath is detected. Once a breath is detected, the device will exit auto pulse delivery.